



CERTIFIED TEST REPORT

CUSTOMER: MCCALLS METER SALES SERVICE
MODEL NO: M0304
METER SERIAL NO: 11-08644

CONFIGURATION

METER INSIDE DIAMETER: 4
METER OUTSIDE DIAMETER: 4.5
TEST DATE: 10/31/2011
TEST FACILITY: Volumetric
IDEAL TEST CONSTANT: 1961

CALIBRATION DATA

	<u>Tested TC</u>	<u>GPM</u>	<u>Accuracy</u>
1	1973	587	100.6

CERTIFIED BY: Paul Hobbs DATE: 11/2/2011

This calibration was performed on a gravimetric or volumetric test facility, traceable to the National Institute of Standards and Technology, USA. The estimated flow measurement uncertainty of the calibration facilities are:
Gravimetric +/- 0.15% Volumetric +/- 0.5%



McCROMETER

3255 WEST STETSON AVENUE
HEMET, CA 92545 USA

PHONE (951) 652-6811 / FAX (951) 652-3078

WEB SITE: <http://www.mccrometer.com> E-MAIL: info@mccrometer.com



11-08644

11/2/2011 5:35:50 PM
Version 1.2 (4/18/2007)